PROTECT AND CARE®

Réf.A00494







COULEURS
Rose - vert clair - blanc - orange - fuchsia

FICHE DE PRODUIT

Gants en nitrile médicaux et HACCP Non stérile

Spécifications

- À usage unique.
- Adapté aux applications HACCP & aux usages médicaux.
- Usage médical & protecteur: prévient les risques de contamination & protège contre les virus, bactéries & champignons.
- 100 % nitrile – Résistance à la traction: 6 N, plus résistant que le latex, le caoutchouc naturel ou le vinyle.
- Flexibles et offrant une sensibilité tactile supérieure aux gants diagnostiques standard.
- Bonne résistance à la perforation et à la déchirure.
- Hypoallergénique : sans latex – sans poudre.
- Imperméable en milieu humide ou gras.
- Ajustement confortable:
  - avec bord roulé pour un enfilage et un retrait faciles
  - ambidextre
- Bouts des doigts microstructurés
- Tailles:

	Longueur mm	Largeur main mm
XS 5-6	**240**	75 +/-10
S 6-7	**240**	80 +/-10
M 7-8	**240**	95 +/-10
L 8-9	**240**	110 +/-10
XL 9-10	**240**	115 +/-10



① min. 0.09 mm
② min. 0.07 mm
③ min. 0.05 mm

- **Conservation:** 5 ans

Conditionnement

	Boîte Distributrice	Carton	Palette
Quantité	100	1000	80 MB
Dimensions	215*120*60mm	315*250*230mm	1.2*0.8*2mt
Poids	525g	5.25kg	537kg

- Emballage & conditionnement pourvus de code QR.
- Notice en 27 langues: NL, FR, EN, DE, ES, IT, PT...

MANUFACTURING AND SAFETY STANDARDS		
AQL	Manufacturing final release: G-I inspection level AQL 1.0 in accordance with ISO 2859-1.	
Latex protein content	Not applicable. Gloves do not contain natural rubber latex.	
Classification	**Medical Device: CE** Class I, Rule 5 (Regulation (EU) 2017/745)	**Personal Protective Equipment: CE 2777** Category III (Regulation (EU) 2016/425)
Product compliances	EN 455-1, EN 455-2, EN 455-3, EN 455-4, EN ISO 15223-1, EN ISO 20417	EN ISO 374-1, EN ISO 374-2, EN ISO 374-4, EN ISO 374-5, EN 16523-1, EN ISO 21420
Quality compliances	EN ISO 13485, ISO 9001	
Viral test permeation	Test in accordance with EN ISO 374-5 (ISO 16604) and ASTM F1671.	
Bacteria and fungi permeation	Test in accordance with EN ISO 374-5 (EN ISO 374-2).	
Chemotherapy drugs permeation test	Test in accordance with ASTM D6978.	
Chemical substances permeation test	Test in accordance with EN 16523-1.	
Biocompatibility / biological evaluation	Test in accordance with EN ISO 10993-5. No cytotoxic evidence observed. Test in accordance with EN ISO 10993-10. No skin irritation and sensitization evidence observed.	
REACH	The product does not contain substances listed in candidate list according to Regulation (EC) 1907/2006.	
Food contact	Gloves intended for food contact based on Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling of food and have been tested for Overall Migration Test acc. EN 1186.	

PROTECTandCARE BV Booiebos,24 9031 Gent **BE** +32 9 321 32 32 www.protectandcare.be