



KLEUREN

Roze - lichtgroen - oranje - fuchsia

PRODUCTFICHE


Medische en HACCP nitril handschoenen Non steriel

Specificaties

- Voor éénmalig gebruik
- Geschikt voor HACCP en medische doeleinden
- Medisch & beschermend: voorkomen besmettingsgevaar en beschermen tegen virussen, bacteriën & fungi
- 100% Nitril - Treksterkte: 6N, sterker dan andere handschoenmaterialen zoals latex, natuur rubber of vinyl.
- Zijn flexibel en bieden een veel grotere tastgevoeligheid dan standaard diagnostische handschoenen.
- Goede punctie- en scheurbestendigheid
- Hypo allergeen: Latexvrij - Poedervrij
- Ondoorlatend bij gebruik in natte of vette werkomgeving
- Comfortabele pasvorm:
 - voorzien van rolrand voor makkelijk aan- en uittrekken
 - zowel links als rechts te dragen
- Microgestructureerde vingertoppen

• Maten:

	Lengte mm	Palmbreedte mm
XS 5-6	240	75 +/-10
S 6-7	240	80 +/-10
M 7-8	240	95 +/-10
L 8-9	240	110 +/-10
XL 9-10	240	115 +/-10



• Houdbaarheid: 5 jaar

Verpakking kwantiteit

- Verpakking voorzien van barcode en QR code
- Info in 27 talen: NL, FR, EN, DE, ES, IT, PT,...

		Masterbox	Pallet
Aantal	100	1000	80 MB
Afmetingen	215*120*60mm	315*250*230mm	1.2*0.8*2mt
Gewicht	525g	5.25kg	448kg

MANUFACTURING AND SAFETY STANDARDS

AQL	Manufacturing final release: G-I inspection level AQL 1.0 in accordance with ISO 2859-1.	
Latex protein content	Not applicable. Gloves do not contain natural rubber latex.	
Classification	Medical Device: CE Class I, Rule 5 (Regulation (EU) 2017/745)	Personal Protective Equipment: CE 2777 Category III (Regulation (EU) 2016/425)
Product compliances	EN 455-1, EN 455-2, EN 455-3, EN 455-4, EN ISO 15223-1, EN ISO 20417	EN ISO 374-1, EN ISO 374-2, EN ISO 374-4, EN ISO 374-5, EN 16523-1, EN ISO 21420
Quality compliances	EN ISO 13485, ISO 9001	
Viral test permeation	Test in accordance with EN ISO 374-5 (ISO 16604) and ASTM F1671.	
Bacteria and fungi permeation	Test in accordance with EN ISO 374-5 (EN ISO 374-2).	
Chemotherapy drugs permeation test	Test in accordance with ASTM D6978.	
Chemical substances permeation test	Test in accordance with EN 16523-1.	
Biocompatibility / biological evaluation	Test in accordance with EN ISO 10993-5. No cytotoxic evidence observed. Test in accordance with EN ISO 10993-10. No skin irritation and sensitization evidence observed.	
REACH	The product does not contain substances listed in candidate list according to Regulation (EC) 1907/2006.	
Food contact	Gloves intended for food contact based on Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling of food and have been tested for Overall Migration Test acc. EN 1186.	